OTS100AF, OTS80AF and OTS60AF Laboratory oil testers



- Laboratory instruments for measuring insulating oil breakdown voltage
- Lock in precision oil vessel with lockable adjustment
- Bright 3.5" colour display visible out doors
- Suitable for mineral, ester and silicon oils
- Trip detection circuit with direct measurement of voltage and current
- Ultra fast (< 10 µs) HV switch off time

DESCRIPTION

Megger's automatic laboratory oil test sets perform accurate breakdown voltage tests on mineral, ester and silicon insulating liquids. Moulded test vessels give repeatable results with lock in precision electrode gap setting adjustment wheels. The transparent, shielded lid and large test chamber enable easy access to the test vessel

All three laboratory models are fitted with a 12 key alpha-numeric keypad to facilitate entry of test ID, filenames, notes etc. Alpha characters are entered by repetitive pressing on a key, the same way as text is entered modern cellular telephones.

Test standards are preloaded in the instrument and new versions can be uploaded via USB flash drive. All laboratory models support the creation of user defined custom tests. Test results are identified either by a serial number or asset ID and are time and date stamped. Megger's asset and data management software, PowerDB Lite, is bundled at no extra cost providing an excellent tool for downloading and printing results.

An optional internal printer provides a hard copy of results. Ink based printout ensures durability at all temperatures. USB interfaces (x3), USB flash drive, external USB printer and barcode scanner.

User safety is paramount and Megger have designed independent and dual redundant high voltage cut-off circuitry to ensure safety. During a test the operator can terminate by pressing any button on the keyboard which will remove high voltage immediately and abort the test. The transparent lid provides ample visibility within the chamber yet is protected and electrically shielded by a screen with multiple links to instrument ground.

APPLICATION

Monitoring and maintenance of oil quality is essential in ensuring the reliable operation of oil filled electrical equipment. Codes of practice have been established in many countries that include several different types of test on insulating oils.

One of the fundamental tests of oil quality is the breakdown voltage test, which is a measure of the oil's ability to withstand electric stress. A low breakdown voltage can indicate the presence of contaminants such as water or conducting particles.

Care should be taken to ensure the process of sampling oil and subsequent testing does not in any way contaminate it with foreign objects. Cleaning vessels between oil tests should be a rinse with the next sample, never clean with fibrous materials. To ensure an accurate reading set gap carefully and lock adjusting wheels.

FEATURES AND BENEFITS

- Test voltages: 60 kV, 80 kV and 100 kV
- Lock in precision oil vessel lockable gap setting
- Flat electrode gap gauges that will not damage electrodes
- Automatic oil temperature measurement
- QVGA colour display with backlight visible in sunlight
- Large, easy clean test chamber with oil drain
- High visibility test chamber
- Safe operation with dual redundant micro switch
- Intuitive user interface

OTS100AF

OPTIONAL FEATURES

■ Internal printer Safety Designed in accordance with IEC61010

■ Motorised lid impeller Dual safety micro switches on chamber ■ Voltage check unit (VCM100D/VCM80D)

EMC Light industrial IEC 61326-1 Class B, **SPECIFICATIONS**

CISPR 22, CISPR 16-1 and CISPR 16-2

English, French, German, Spanish,

Test voltage Dimensions All models 580 mm x 420 mm x 290 mm OTS60AF 0 to 60 kV rms maximum (30 kV - 0 - 30 kV)

Weight All models 30 kg with printer option fitted

OTS80AF 0 to 80 kV rms maximum (40 kV - 0 - 40 kV) 0 to 100 kV rms maximum (50 kV - 0 - 50 kV)

Language

Voltage resolution 0.1 kV °1% °2 digits Chinese, Czech, Dutch, Finnish, Italian, and accuracy Norwegian, Polish, Portuguese, Russian **Programmed test sequences**

and Swedish

ASTM D 1816-04 BS EN 60156-96 SABS EN60156

ASTM D 877A-02 CEI EN 60156-95 VDE0370 part 5 ASTM D 877B-02 **IRAM 2341** AS1767.2.1 IEC 60156-95 UNE EN 60156 PA SEV EN60156

NF EN 60156 JIS C 2101-99 (M)

JIS C 2101-99 (S) plus 3 custom test sequences

400 ml (standard) 150 ml (option) Vessels

Nylon 12 chamber provides precision electrode alignment, adjustment wheels lock electrodes in position, option of 150 ml vessel for low volume oil samples

10 °C to 65 °C **Temperature**

measuring range

Temperature sensor resolution

1 °C

Line voltage 85 to 265 VAC **Power supply**

Line frequency 50/60 Hz

Interface USB 2.0 compatible

2 x USB type-A (Flash drive, printer, other)

1 x USB type-B (Factory use only)

Internal printer

Matrix impact printer Paper 57.5 mm wide

External printer Any printer with USB interface and PCL3

driver

Protection Dual safety micro switches on chamber

Display 320 x 240 QVGA colour display with

backlight

Operating temperature range and humidity

0 °C to +50 °C 80% RH at 40 °C

Storage temperature -30 °C to +65 °C range and humidity 95% RH at 40 °C

Maximum altitude 2000 m

Programmed test sequence overview

	ے																
Breakdown test sequence	Time between tests	2 mins	2 mins	2 mins	2 mins	2 mins	2 mins	2 mins	2 mins	2 mins	2 mins	1 min	N/A	1 min 15s	1 min	N/A	10s to 600s
	Intial stand time	5 mins	5 mins	5 mins	5 mins	5 mins	5 mins	5 mins	5 mins	5 mins	5 mins	2 mins	2 mins (x5)	3 mins	2 mins	2 mins (x5)	10s to 600s
	Number of tests	9	9	9	9	9	9	9	9	9	9	5 x 2	1 x 5	2	5	1 x 5	5, 6 or 10
otions	3 kV/s											•	•		•	•	0.5 kV/s to 5 kV/s
Voltage rise rate options	2 kV/s	-	•	•	•	•	•	-	-	-	-						
Voltage	0.5 kV/s													•			0.5 4
Oil stirring options		-	•	•	•	•	•	-	•	•	-	•					•
	•	•	•	•	•	•	•	•	•	•	•	•		•			•
	X	-	•		•	•		-	•	•	•		•		•	•	•
options	8														-	•	•
Electrode shape options		•	•		•	•				•		•	•				•
Electro	-)-	•	•	•	•	•	•	•	•	•	•			•			•
Electrode gap options (mm)	2.54														•		1.0 to 7.0
	2.5		•		•	•		•			•	•	•				
:lectro	2.0													•			1.0
	1.0																
s tested	Silicon	•	•	•	•	•	•	•		•	•		•	•	•	•	•
Oil types tested	Mineral Ester HMWH	•	•	•	•		•	•			•	•		•	•	•	•
	Standards complied with and programmed	IEC 60156-95	BS EN 60156-96	CEI EN 60156-95	IRAM 2341	UNI EN 60156	NF EN 60156	SABS EN 60156	VDE 0370 part 5	AS1767.2.1	PA SEV EN 60156	JIS C 2101-99 (M)	JIS C 2101-99 (S)	ASTM D 1816-04	ASTM D 877A-02	ASTM D 877B-02	Custom tests (x3) (Programmable)

ORDERING CONFIGURATION

Example of an ordering configuration:-

OTS100AF-USA-P4 = This order is for an OTS100AF with US power lead, ASTM electrode set, internal printer and lid stirrer.

Model:	OTS		AF-		-		
							Weights
	100 kV	100					29.5 kg
Select Model	80 kV	80					29.5 kg
	60 kV	60					29.5 kg
		EU Lead	EU				
Select Power	m [m	UK Lead	UK				
Cord		US Lead	US				
		Australian Lead	AU				
		No Plug Lead	BL				
	4++	f	ASTM Set	A			
Select Electrode Set	•••	•	IEC Set	E			
)(41	Full Set	U			
Select Printer			BE BE	Internal P Printer			0.54 kg
Option		Megger orssers		No Printer	х		0.08 kg
Select Stirrer			а	•	Stirrer Lid Assembly Fitted	4	0.3 kg
Option			Stirrer Lid Assembly Not Fitted	x			

Description	Order Code	Description	Order Code				
configured*	OTS60AF	36 mm mushroom electrodes (2)					
configured*	OTS80AF	Magnetic stirrer bar (2)					
configured*	OTS100AF	Magnetic stirrer bar retriever (1)					
Included accessories (on all configuration	ons)	Gap gauge set	1001-478				
Vessel 400 ml assembly		Full electrode set (covers IEC and ASTM star	dards)				
Magnetic bead stirrers (2 off)		12.7 mm spherical electrodes (2)					
Magnetic bead retriever		36 mm mushroom electrodes (2)					
User manual CD		25.4 mm standard (sharp edges) cylindrical elect	rodes (2)				
PowerDB Lite software		25.4 mm non-standard (round edges) cylindrical	electrodes (2)				
Electrode gauge set 1, 2, 2.5, 2.54 mm	1002-144	Magnetic stirrer bar (2)					
Configured accessories (to order addition	onal or spares)	Magnetic stirrer bar retriever (1)					
OTS IEC60156 Electrode set contents	mar or spares,	Gap gauge set	1001-479				
- supplied in accessory case		Vessel lid mounted impeller (ASTM D1816) for use with 400 ml - vessel 1001-102					
12.7 mm spherical electrodes (2)		Optional accessories					
36 mm mushroom electrodes (2)		Vessel 400 ml assembly (no electrodes supplied)	1001-473				
Magnetic stirrer bar (2)		Vessel 150 ml assembly (no electrodes supplied)	1001-473				
Magnetic stirrer bar retriever (1)		VCM100D digital voltage checker	1001-474				
Gap gauge set		VCM80D digital voltage checker	1001-103				
OTS ASTM D877/D1816 Electrode set co	1001-477 ntents	Printer paper, 1 roll (MOV applies) (4 rolls supplied if printer configured) 25995-001					
- supplied in accessory case	.11 (2)	Barcode reader, USB	1001-047				
25.4 mm standard (sharp edges) cylindrical		NOTE:					
25.4 mm non-standard (round edges) cylind	irical electrodes (2)	* See ordering configuration on previous page					
		See ordering configuration on previous page					

UK Archcliffe Road Dover CT17 9EN England T +44 (0) 1304 502101 F +44 (0) 1304 207342 UKsales@megger.com

UNITED STATES

UNITED STATES
4271 Bronze Way
Dallas TX 75237-1019 USA
T 800 723 2861 (USA only)
T +1 214 333 3201
F +1 214 331 7399 USsales@megger.com

OTHER TECHNICAL SALES OFFICES
Valley Forge USA, College Station USA,
Sydney AUSTRALIA, Danderyd SWEDEN,
Ontario CANADA, Trappes FRANCE,
Oberursel GERMANY, Aargau SWITZERLAND,
Kingdom of BAHRAIN, Mumbai INDIA,
Johannesburg SOUTH AFRICA, Chonburi THAILAND

CERTIFICATION ISO

Registered to ISO 9001:2008 Cert. no. Q 09290 Registered to ISO 14001-2004 Cert. no. EMS 61597 OTS100AF--OTS80AF--OTS60AF _DS_en_V08.pdf